



This report cannot be used for ODA, OHA or OLCC compliance requirements.

Product identity: 200mg Strawberry Milk
Laboratory ID: 19-010832-0001

Client/Metric ID: .
Sample Date: 09/04/19 11:00

Summary

Potency:

Analyte	Result (%)		
CBD	39.5		CBD-Total 39.5%
			THC-Total < 0.165% (Reported in percent of total sample)

Pesticides:

All analytes passing and less than LOQ.



This report cannot be used for ODA, OHA or OLCC compliance requirements.

Customer: Hempbase
200 Spectrum Center Dr FL 3
Irvine, CA 92618
United States

Product identity: 200mg Strawberry Milk

Client/Metric ID: .


Sample Date: 09/04/19 11:00

Laboratory ID: 19-010832-0001

Relinquished by: Received By Mail

Temp: 24.2 °C

Sample Results

Potency	Method J AOAC 2015 V98-6			Units %	Batch 1908309	Analyze 09/12/19 12:08 AM
Analyte	As Received	Dry weight	LOQ	Notes	 <ul style="list-style-type: none"> ● CBD 	
CBC†	< LOQ		0.0881			
CBC-A†	< LOQ		0.0881			
CBC-Total†	< LOQ		0.165			
CBD	39.5		0.881			
CBD-A	< LOQ		0.0881			
CBD-Total	39.5		0.958			
CBDV†	< LOQ		0.0881			
CBDV-A†	< LOQ		0.0881			
CBDV-Total†	< LOQ		0.164			
CBG†	< LOQ		0.0881			
CBG-A†	< LOQ		0.0881			
CBG-Total†	< LOQ		0.164			
CBL†	< LOQ		0.0881			
CBN	< LOQ		0.0881			
Δ8-THC†	< LOQ		0.0881			
Δ9-THC	< LOQ		0.0881			
THC-A	< LOQ		0.0881			
THC-Total	< LOQ		0.165			
THCV†	< LOQ		0.0881			
THCV-A†	< LOQ		0.0881			
THCV-Total†	< LOQ		0.164			



This report cannot be used for ODA, OHA or OLCC compliance requirements.

Pesticides											Method AOAC 2007.01 & EN 15662 (mod)					Units mg/kg		Batch 1908239		Analyze 09/13/19 11:42 AM				
Analyte	Result	Limits	LOQ	Status	Notes	Analyte	Result	Limits	LOQ	Status	Notes	Analyte	Result	Limits	LOQ	Status	Notes							
Abamectin	< LOQ	0.50	0.250	pass		Acephate	< LOQ	0.40	0.250	pass														
Acequinocyl	< LOQ	2.0	1.00	pass		Acetamiprid	< LOQ	0.20	0.100	pass														
Aldicarb	< LOQ	0.40	0.200	pass		Azoxystrobin	< LOQ	0.20	0.100	pass														
Bifenazate	< LOQ	0.20	0.100	pass		Bifenthrin	< LOQ	0.20	0.100	pass														
Boscalid	< LOQ	0.40	0.100	pass		Carbaryl	< LOQ	0.20	0.100	pass														
Carbofuran	< LOQ	0.20	0.100	pass		Chlorantraniliprole	< LOQ	0.20	0.100	pass														
Chlorfenapyr	< LOQ	1.0	0.500	pass		Chlorpyrifos	< LOQ	0.20	0.100	pass														
Clofentezine	< LOQ	0.20	0.100	pass		Cyfluthrin (incl.	< LOQ	1.0	0.500	pass														
Cypermethrin	< LOQ	1.0	0.500	pass		Daminozide	< LOQ	1.0	0.500	pass														
Diazinon	< LOQ	0.20	0.100	pass		Dichlorvos	< LOQ	1.0	0.500	pass														
Dimethoate	< LOQ	0.20	0.100	pass		Ethoprophos	< LOQ	0.20	0.100	pass														
Etofenprox	< LOQ	0.40	0.200	pass		Etoxazole	< LOQ	0.20	0.100	pass														
Fenoxycarb	< LOQ	0.20	0.100	pass		Fenpyroximate	< LOQ	0.40	0.200	pass														
Fipronil	< LOQ	0.40	0.200	pass		Flonicamid	< LOQ	1.0	0.400	pass														
Fludioxonil	< LOQ	0.40	0.200	pass		Hexythiazox	< LOQ	1.0	0.400	pass														
Imazalil	< LOQ	0.20	0.100	pass		Imidacloprid	< LOQ	0.40	0.200	pass														
Kresoxim-methyl	< LOQ	0.40	0.200	pass		Malathion	< LOQ	0.20	0.100	pass														
Metalaxyl	< LOQ	0.20	0.100	pass		Methiocarb	< LOQ	0.20	0.100	pass														
Methomyl	< LOQ	0.40	0.200	pass		MGK-264	< LOQ	0.20	0.100	pass														
Myclobutanil	< LOQ	0.20	0.100	pass		Naled	< LOQ	0.50	0.250	pass														
Oxamyl	< LOQ	1.0	0.500	pass		Paclobutrazole	< LOQ	0.40	0.200	pass														
Parathion-Methyl	< LOQ	0.20	0.200	pass		Permethrin	< LOQ	0.20	0.100	pass														
Phosmet	< LOQ	0.20	0.100	pass		Piperonyl butoxide	< LOQ	2.0	1.00	pass														
Prallethrin	< LOQ	0.20	0.100	pass		Propiconazole	< LOQ	0.40	0.200	pass														
Propoxur	< LOQ	0.20	0.100	pass		Pyrethrin I (total)	< LOQ	1.0	0.500	pass														
Pyridaben	< LOQ	0.20	0.100	pass		Spinosad	< LOQ	0.20	0.100	pass														
Spiromesifen	< LOQ	0.20	0.100	pass		Spirotetramat	< LOQ	0.20	0.100	pass														
Spiroxamine	< LOQ	0.40	0.200	pass		Tebuconazole	< LOQ	0.40	0.200	pass														
Thiacloprid	< LOQ	0.20	0.100	pass		Thiamethoxam	< LOQ	0.20	0.100	pass														
Trifloxystrobin	< LOQ	0.20	0.100	pass																				



This report cannot be used for ODA, OHA or OLCC compliance requirements.

Abbreviations

Limits: Action Levels per OAR-333-007-0400, OAR-333-007-0210, OAR-333-007-0220

Limit(s) of Quantitation (LOQ): The minimum levels, concentrations, or quantities of a target variable (e.g., target analyte) that can be reported with a specified degree of confidence.

† = Analyte not NELAP accredited.

Units of Measure

mg/kg = Milligram per kilogram = parts per million (ppm)

% = Percentage of sample

% wt = µg/g divided by 10,000

Approved Signatory

Derrick Tanner
General Manager



This report cannot be used for ODA, OHA or OLCC compliance requirements.

Revision: 1.00 Control: CFL-C2
Revised:08/12/2019 Effective: 08/15/2019

Laboratory Pesticide Quality Control Results

AOAC2007.1 & EN 15662		Units: mg/Kg		Batch ID 1908239				
Method Bank	Laboratory Control Sample							
Analyte	Blank Result	Blank Limits	Notes	LCS Result	LCS Spk	LCS % Re	Limits	Notes
Acophate	0.00	< 0.200		0.990	1.000	99.0	74.4 - 130	
Acetaminophen	0.00	< 1.000		4.382	4.000	109.6	87.7 - 118	
Acetaminoprid	0.000	< 0.100		0.419	0.400	104.6	90.1 - 115	
Aldicarb	0.000	< 0.200		0.801	0.800	100.1	88.7 - 116	
Abamectin	0.000	< 0.288		1.023	1.000	102.3	82.6 - 122	
Azoxystrobin	0.000	< 0.100		0.431	0.400	107.7	83.4 - 125	
Bifenazate	0.000	< 0.100		0.415	0.400	103.8	89.1 - 118	
Bifenthrin	0.001	< 0.100		0.379	0.400	94.6	82.5 - 119	
Boscalid	0.000	< 0.100		0.777	0.800	97.1	78.9 - 127	
Carbaryl	0.000	< 0.100		0.415	0.400	103.8	86.7 - 120	
Carbofuran	0.000	< 0.100		0.417	0.400	104.2	88.9 - 121	
Chlorantraniliprol	0.000	< 0.100		0.418	0.400	104.6	69.7 - 128	
Chlorfenapyr	0.000	< 1.000		1.952	2.000	97.6	72.0 - 131	
Chlorpyrifos	0.000	< 0.100		0.474	0.400	118.6	84.2 - 123	
Clofentezine	0.000	< 0.100		0.380	0.400	97.5	84.0 - 115	
Cyfluthrin	0.003	< 1.000		1.983	2.000	99.1	83.2 - 120	
Cypermethrin	0.000	< 1.000		2.210	2.000	110.5	89.1 - 118	
Daminozide	0.000	< 1.000		1.973	2.000	98.7	81.0 - 119	
Daaznon	0.000	< 0.100		0.437	0.400	109.1	87.8 - 118	
Dichlorvos	0.000	< 0.500		1.881	2.000	94.0	82.1 - 117	
Dimethoat	0.000	< 0.100		0.418	0.400	104.4	88.5 - 116	
Ethoprophos	0.000	< 0.100		0.401	0.400	100.2	83.5 - 120	
Etofenprox	0.000	< 0.100		0.884	0.800	108.0	90.2 - 123	
Etoxazol	0.000	< 0.100		0.383	0.400	95.8	85.5 - 120	
Fenoxycarb	0.000	< 0.100		0.416	0.400	104.0	84.9 - 121	
Fenpyroximat	0.000	< 0.100		0.851	0.800	106.4	87.1 - 124	
Flpronil	0.000	< 0.100		0.854	0.800	106.8	81.9 - 126	
Flonicamid	0.001	< 0.400		1.066	1.000	106.6	82.2 - 119	
Fludioxonil	0.000	< 0.100		0.839	0.800	104.9	87.4 - 128	
Hexythiazox	0.000	< 0.400		1.066	1.000	106.6	90.6 - 120	
Imazalil	0.000	< 0.100		0.405	0.400	101.3	87.5 - 124	
Imidacloprid	0.000	< 0.200		0.811	0.800	101.4	84.5 - 119	
Kiesoxim-Methyl	0.000	< 0.100		0.840	0.800	105.0	91.2 - 115	
Malathion	0.000	< 0.100		0.343	0.400	85.8	85.8 - 122	
Metaxalyl	0.000	< 0.100		0.416	0.400	104.0	87.3 - 119	
Methiocarb	0.000	< 0.100		0.395	0.400	98.8	81.4 - 122	
Methomyl	0.000	< 0.200		0.809	0.800	101.2	76.1 - 120	
MCK 264	0.000	< 0.100		0.405	0.400	101.3	87.3 - 120	
Mydobutanol	0.000	< 0.100		0.443	0.400	110.6	89.9 - 116	
Naled	0.000	< 0.200		1.086	1.000	108.6	87.6 - 122	
Oxamyl	0.000	< 0.400		2.030	2.000	101.5	80.8 - 117	
Padobutrazol	0.000	< 0.200		0.840	0.800	105.1	87.0 - 122	
Parathion Methyl	0.000	< 0.200		0.852	0.800	106.5	72.4 - 134	
Permethrin	0.000	< 0.100		0.462	0.400	115.4	86.3 - 120	
Phosmet	0.000	< 0.100		0.416	0.400	104.0	90.3 - 117	
Piperonyl butoxide	0.000	< 1.000		2.336	2.000	116.8	88.3 - 133	
Prallethrin	0.000	< 0.200		0.406	0.400	101.4	89.0 - 120	
Propiconazole	0.000	< 0.200		0.867	0.800	108.4	92.2 - 113	
Propoxur	0.000	< 0.100		0.405	0.400	101.4	86.6 - 120	
Pyrethrins	0.000	< 0.500		0.253	0.284	88.9	79.5 - 140	
Pyridaben	0.000	< 0.100		0.473	0.400	118.2	83.0 - 138	
Spinosad	0.000	< 0.100		0.411	0.388	105.9	94.4 - 122	
Spiromesifen	0.005	< 0.100		0.431	0.400	107.8	85.3 - 120	
Spirotetramat	0.000	< 0.100		0.419	0.400	104.8	80.9 - 123	
Spiroxamine	0.000	< 0.100		0.840	0.800	105.0	86.3 - 131	
Tebuconazol	0.000	< 0.200		0.809	0.800	101.2	87.6 - 120	
Thiadoprid	0.000	< 0.100		0.402	0.400	100.6	89.6 - 119	
Thiamethoxam	0.000	< 0.100		0.397	0.400	99.2	80.9 - 122	
Trifloxystrobin	0.000	< 0.100		0.425	0.400	106.4	90.0 - 118	

Test results relate only to the parameters tested and to the samples as received by the laboratory. Test results meet all requirements of NELAP and the Pixis quality assurance plan unless otherwise noted. This report shall not be reproduced, except in full, without the written consent of this laboratory. Samples will be kept a maximum of 15 days from the report date unless prior arrangements have been made.



This report cannot be used for ODA, OHA or OLCC compliance requirements.

Revision: 1.00 Control: CFL-C2
Revised: 08/12/2019 Effective: 08/15/2019

Laboratory Pesticide Quality Control Results

AOAC2007.1 & EN 15662		Units: mg/Kg					Batch ID 1908239				
Matrix Spike/Matrix Spike Duplicate Recoveries		Sample ID: 19-0108100002									
Analyte	Result	MSR _{as}	MSD _{as}	Spike	RFD%	Limit	MS% Re	MSD% Re	Limits	Notes	
Acephate	0.00	1.037	1.122	1.000	7.9	< 30	103.7	112.2	50 - 150		
Acequinocyl	0.00	4.874	4.884	4.000	0.2	< 30	121.9	122.1	50 - 150		
Acetamiprid	0.00	0.412	0.419	0.400	1.7	< 30	103.0	104.8	50 - 150		
Aldicarb	0.00	0.788	0.808	0.800	2.5	< 30	98.5	101.0	50 - 150		
Abamectin	0.00	1.100	1.033	1.000	6.2	< 30	110.0	103.3	50 - 150		
Azoxystrobin	0.00	0.475	0.486	0.400	2.4	< 30	118.7	121.6	50 - 150		
Bifenazate	0.00	0.437	0.443	0.400	1.3	< 30	109.3	110.7	50 - 150		
Bifenthrin	0.00	0.970	0.905	0.400	6.9	< 30	242.4	226.3	50 - 150	Q1	
Boscalid	0.00	0.807	0.955	0.800	16.8	< 30	100.8	119.4	50 - 150		
Carbaryl	0.00	0.417	0.433	0.400	3.8	< 30	104.3	108.3	50 - 150		
Carbofuran	0.00	0.449	0.411	0.400	8.8	< 30	112.2	102.7	50 - 150		
Chlorantraniliprol	0.00	0.378	0.433	0.400	13.6	< 30	94.5	108.3	50 - 150		
Chlorfenapyr	0.00	2.233	2.244	2.000	0.5	< 30	111.6	112.2	50 - 150		
Chlorpyrifos	0.00	0.537	0.484	0.400	10.4	< 30	134.3	121.0	50 - 150		
Clofentezane	0.00	0.447	0.443	0.400	0.7	< 30	111.7	110.8	50 - 150		
Cyfluthrin	0.00	1.997	1.728	2.000	14.5	< 30	99.9	86.4	30 - 150		
Cypermethrin	0.00	3.772	3.833	2.000	1.6	< 30	188.6	191.6	50 - 150	Q1	
Daminozide	0.00	1.806	1.782	2.000	1.4	< 30	90.3	89.1	30 - 150		
Diazinon	0.00	0.419	0.402	0.400	4.1	< 30	104.7	100.5	50 - 150		
Dichlorvos	0.00	2.000	1.948	2.000	2.6	< 30	100.0	97.4	50 - 150		
Dimethoat	0.00	0.441	0.418	0.400	5.2	< 30	110.2	104.6	50 - 150		
Ethoprophos	0.00	0.402	0.398	0.400	0.9	< 30	100.4	99.5	50 - 150		
Etofenprox	0.00	0.867	0.866	0.800	0.1	< 30	108.4	108.2	50 - 150		
Etoxazol	0.00	0.143	0.168	0.400	16.0	< 30	35.8	42.0	50 - 150	Q	
Fenoxycarb	0.00	0.424	0.428	0.400	0.9	< 30	108.1	107.0	50 - 150		
Fenpyroximat	0.00	0.335	0.377	0.800	6.3	< 30	116.8	109.6	50 - 150		
Fipronil	0.00	0.912	0.980	0.800	7.2	< 30	114.0	122.5	50 - 150		
Fonicamid	0.00	0.828	1.022	1.000	21.0	< 30	82.8	102.2	50 - 150		
Fludioxonil	0.00	0.887	0.895	0.800	0.9	< 30	110.9	111.8	50 - 150		
Hexythiazox	0.00	1.248	1.205	1.000	3.5	< 30	124.8	120.5	50 - 150		
Imazalil	0.00	0.413	0.428	0.400	3.8	< 30	103.2	107.1	50 - 150		
Imidacloprid	0.00	0.854	0.859	0.800	0.6	< 30	106.7	107.4	50 - 150		
Kiesoxim-Methyl	0.00	0.888	0.889	0.800	0.1	< 30	110.9	111.1	50 - 150		
Malathion	0.00	0.385	0.343	0.400	11.6	< 30	96.2	85.7	50 - 150		
Metaxalyl	0.00	0.465	0.445	0.400	4.4	< 30	116.3	111.3	50 - 150		
Methiocarb	0.00	0.427	0.432	0.400	1.1	< 30	106.9	108.1	50 - 150		
Methomyl	0.00	0.826	0.859	0.800	4.0	< 30	103.2	107.4	50 - 150		
MCK 264	0.00	0.500	0.456	0.400	9.3	< 30	125.1	114.0	50 - 150		
Mydobutanol	0.00	0.424	0.410	0.400	3.4	< 30	105.9	102.4	50 - 150		
Naled	0.00	1.190	1.194	1.000	0.3	< 30	119.0	119.4	50 - 150		
Oxamyl	0.00	2.085	2.065	2.000	1.0	< 30	104.3	103.2	50 - 150		
Padobutrazol	0.00	0.897	0.926	0.800	3.2	< 30	112.1	115.8	50 - 150		
Parathion Methyl	0.00	0.643	0.782	0.800	19.5	< 30	80.4	97.8	30 - 150		
Permethrin	0.00	0.532	0.537	0.400	1.1	< 30	132.9	134.4	50 - 150		
Phosmet	0.00	0.447	0.459	0.400	2.7	< 30	111.7	114.7	50 - 150		
Piperonyl butoxide	0.00	2.535	2.569	2.000	1.3	< 30	126.8	128.5	50 - 150		
Prallethrin	0.00	0.577	0.598	0.400	3.6	< 30	144.2	149.5	50 - 150		
Propiconazole	0.00	0.891	0.902	0.800	1.3	< 30	111.4	112.8	50 - 150		
Propoxur	0.00	0.417	0.412	0.400	1.2	< 30	104.2	102.9	50 - 150		
Pyrethrins	0.00	0.210	0.210	0.284	0.2	< 30	73.8	73.9	50 - 150		
Pyridaben	0.00	0.507	0.496	0.400	2.1	< 30	126.6	124.0	50 - 150		
Spinosad	0.00	0.455	0.442	0.388	3.0	< 30	117.3	113.9	50 - 150		
Spiromesifen	0.00	0.546	0.528	0.400	3.2	< 30	136.3	132.0	50 - 150		
Spirotetramat	0.00	0.355	0.374	0.400	5.2	< 30	88.7	93.5	50 - 150		
Spiroxamine	0.00	0.829	0.881	0.800	6.1	< 30	103.6	110.1	50 - 150		
Tebuconazol	0.00	0.799	0.861	0.800	7.5	< 30	99.9	107.6	50 - 150		
Thiadoprid	0.00	0.410	0.442	0.400	7.4	< 30	102.5	110.4	50 - 150		
Thiamethoxam	0.00	0.398	0.425	0.400	6.7	< 30	99.5	106.3	50 - 150		
Trifloxystrobin	0.00	0.441	0.451	0.400	1.5	< 30	110.2	112.7	50 - 150		

Test results relate only to the parameters tested and to the samples as received by the laboratory. Test results meet all requirements of NELAP and the Pixis quality assurance plan unless otherwise noted. This report shall not be reproduced, except in full, without the written consent of this laboratory. Samples will be kept a maximum of 15 days from the report date unless prior arrangements have been made.



This report cannot be used for ODA, OHA or OLCC compliance requirements.

Laboratory Quality Control Results

JAOAC2015 V98-6		Batch ID: 1908309							
Laboratory Control Sample									
Analyte	Result	Spike	Units	% Rec	Limits	Evaluation	Notes		
CBDV-A	0.186	0.2	%	93.0	85 - 115	Acceptable			
CBDV	0.187	0.2	%	93.5	85 - 115	Acceptable			
CBD-A	0.197	0.2	%	98.5	85 - 115	Acceptable			
CBG-A	0.189	0.2	%	94.5	85 - 115	Acceptable			
CBG	0.193	0.2	%	96.5	85 - 115	Acceptable			
CBD	0.202	0.2	%	101	85 - 115	Acceptable			
THCV	0.194	0.2	%	97.0	85 - 115	Acceptable			
THCVA	0.183	0.2	%	91.5	85 - 115	Acceptable			
CBN	0.192	0.2	%	96.0	85 - 115	Acceptable			
THC	0.196	0.2	%	98.0	85 - 115	Acceptable			
D8THC	0.184	0.2	%	92.0	85 - 115	Acceptable			
CBL	0.180	0.2	%	90.0	85 - 115	Acceptable			
CBC	0.187	0.2	%	93.5	85 - 115	Acceptable			
THCA	0.185	0.2	%	92.5	85 - 115	Acceptable			
CBCA	0.181	0.2	%	90.5	85 - 115	Acceptable			

Method Blank

Analyte	Result	LOQ	Units	Limits	Evaluation	Notes	
CBDV-A	ND	0.1	%	< 0.1	Acceptable		
CBDV	ND	0.1	%	< 0.1	Acceptable		
CBD-A	ND	0.1	%	< 0.1	Acceptable		
CBG-A	ND	0.1	%	< 0.1	Acceptable		
CBG	ND	0.1	%	< 0.1	Acceptable		
CBD	ND	0.1	%	< 0.1	Acceptable		
THCV	ND	0.1	%	< 0.1	Acceptable		
THCVA	ND	0.1	%	< 0.1	Acceptable		
CBN	ND	0.1	%	< 0.1	Acceptable		
THC	ND	0.1	%	< 0.1	Acceptable		
D8THC	ND	0.1	%	< 0.1	Acceptable		
CBL	ND	0.1	%	< 0.1	Acceptable		
CBC	ND	0.1	%	< 0.1	Acceptable		
THCA	ND	0.1	%	< 0.1	Acceptable		
CBCA	ND	0.1	%	< 0.1	Acceptable		

Abbreviations

ND - None Detected at or above MRL
RPD - Relative Percent Difference
LOQ - Limit of Quantitation

Units of Measure:

% - Percent



This report cannot be used for ODA, OHA or OLCC compliance requirements.

JAOAC2015 V986		Batch ID: 1908309						
Sample Duplicate		Sample ID: 19-010798-0001						
Analyte	Result	Org. Result	LOQ	Units	RPD	Limits	Evaluation	Notes
CBDV-A	ND	ND	0.1	%	0	< 20	Acceptable	
CBDV	0.209	0.211	0.1	%	0.952	< 20	Acceptable	
CBD-A	ND	ND	0.1	%	0	< 20	Acceptable	
CBG-A	ND	ND	0.1	%	0	< 20	Acceptable	
CBG	ND	ND	0.1	%	0	< 20	Acceptable	
CBD	26.4	26.3	0.1	%	0.380	< 20	Acceptable	
THCV	ND	ND	0.1	%	0	< 20	Acceptable	
THCVA	ND	ND	0.1	%	0	< 20	Acceptable	
CBN	ND	ND	0.1	%	0	< 20	Acceptable	
THC	ND	ND	0.1	%	0	< 20	Acceptable	
D8THC	ND	ND	0.1	%	0	< 20	Acceptable	
CBL	ND	ND	0.1	%	0	< 20	Acceptable	
CBC	ND	ND	0.1	%	0	< 20	Acceptable	
THCA	ND	ND	0.1	%	0	< 20	Acceptable	
CBCA	ND	ND	0.1	%	0	< 20	Acceptable	

Abbreviations

ND - None Detected at or above MRL
RPD - Relative Percent Difference
LOQ - Limit of Quantitation

Units of Measure:

% - Percent



This report cannot be used for ODA, OHA or OLCC compliance requirements.

Explanation of QC Flag Comments:

Code	Explanation
Q	Matrix interferences affecting spike or surrogate recoveries.
Q1	Quality control result biased high. Only non-detect samples reported.
Q2	Quality control outside QC limits. Data considered estimate.
Q3	Sample concentration greater than four times the amount spiked.
Q4	Non-homogenous sample matrix, affecting RPD result and/or % recoveries.
Q5	Spike results above calibration curve.
Q6	Quality control outside QC limits. Data acceptable based on remaining QC.
R	Relative percent difference (RPD) outside control limit.
R1	RPD non-calculable, as sample or duplicate results are less than five times the LOQ.
R2	Sample replicates RPD non-calculable, as only one replicate is within the analytical range.
LOQ1	Quantitation level raised due to low sample volume and/or dilution.
LOQ2	Quantitaion level raised due to matrix interference.
B	Analyte detected in method blank, but not in associated samples.
B1	The sample concentration is greater than 5 times the blank concentration.
B2	The sample concentration is less than 5 times the blank concentration.